



## Certificate of Analysis

### CDI Labs HuProt™ Human Proteome Microarray v3.1

**Product Number:** CDIHP-003.1

**Lot Number:** HuProt\_v3.1\_JAN23\_2018

**Date of Manufacture:** January 23, 2017

**Expiration Date:** 12 months for most applications (6 months for functional assays)

**Surface:** PATH

**Storage conditions:** Store in plastic slide holders at -80°C (slide holders containing microarrays may be stored at -80°C in a plastic bag that contains desiccant).

**Uses:** Research use only. Do not use in diagnostic procedures.

---

#### Product Background:

The **HuProt™** v3.1 Human Proteome Microarray is comprised of >19,000 unique human proteins that allow hundreds of interactions in high-throughput to be profiled. **HuProt™** v3.1 can be used for a wide range of applications including the study of antibody specificity determination, serum profiling, protein-protein interaction, substrate identification, protein-DNA binding, protein-RNA binding, etc.

---

#### Printing Consistency:

>95% of the spots showed a foreground/background signal (F/B) ratio of at least 1.5 in anti-GST assay.

**Result:** Meets all specifications

#### Controls:

The controls on the array are reactive with secondary detection reagents. Each block contains a row of control spots, including Alex Fluor 555/647 as landmarks. Controls include titrated GST protein, histones, mouse and rabbit anti-biotin, mouse IgM, and biotin-tagged control for streptavidin detection.

**Result:** This batch meets the specifications as listed above.

If you have any further questions about this Certificate of Analysis, please contact CDI Technical Services at 787-806-4100 ext 233, or at [info@cdi-lab.com](mailto:info@cdi-lab.com).

Quality Assurance Issued on **January 31, 2017**

CDI Laboratories  
Guanajibo Research and Innovation Park 1300011776  
4005 Street B, Road 114 Km 1.3  
Mayaguez, PR 00682