



## Certificate of Analysis

### CDI Labs HuProt™ Human Proteome Microarray v4.0

**Product Number:** CDIHP-004.0

**Lot Number:** HuProt\_v4.0\_NOV26\_2019

**Date of Manufacture:** November 26, 2018

**Expiration Date:** 12 months for most applications (6 months for functional assays)

**Surface:** SuperEpoxy2

**Storage conditions:** Store in plastic slide holders at -80°C (slide holders containing microarrays may be stored at -80°C in a plastic bag that contains desiccant).

**Uses:** Research use only. Do not use in diagnostic procedures.

---

#### Product Background:

The **HuProt™** v4.0 Human Proteome Microarray is comprised of over 21,000 human proteins and protein isoforms, including >81% of canonically expressed proteins as defined by the Human Protein Atlas, that allow hundreds of interactions in high-throughput to be profiled. HuProt™ v4.0 can be used for a wide range of applications including the study of antibody specificity determination, serum profiling, protein-protein interaction, substrate identification, protein-DNA binding, protein-RNA binding, etc.

---

#### Printing Consistency:

>95% of the spots showed a foreground/background signal (F/B) ratio of at least 1.5 in anti-GST assay.

**Result:** Meets all specifications

#### Controls:

The controls on the array are reactive with secondary detection reagents. Each block contains a row of control spots, including Alex Fluor 555/647 as landmarks. Controls include titrated GST protein, histones, mouse and rabbit anti-biotin, mouse IgM, and biotin-tagged control for streptavidin detection.

**Result:** This batch meets the specifications as listed above.

If you have any further questions about this Certificate of Analysis, please contact CDI Technical Services at 787-806-4100 ext 233, or at [info@cdi-lab.com](mailto:info@cdi-lab.com).

Quality Assurance was issued on **December 06, 2018**

CDI Laboratories  
12 W. Mendez Vigo, 2<sup>nd</sup> Floor  
Mayaguez, PR 00680  
USA