

User Guide

HuProt™ Human Proteome Microarray v3.1

CDI Laboratories
P.O. Box 569
Mayaguez, PR 00681-0569

939-280-5293
844-539-6296 (Toll Free)
support@cdi-lab.com
www.cdi-lab.com

TABLE OF CONTENTS

SUMMARY	1
Storage and Handling	3
Overview - Key Steps for HuProt™ Microarray Use	4
Additional Reagents and Materials	6
<hr/>	
ASSAYS	
Antibody Specificity	8
Antibody Specificity (Using Denatured Arrays)	11
Serum Profiling	14
RNA Binding	17
DNA Binding	19
Protein Binding	21
<hr/>	
APPENDIX	
Recipes	24
HuProt™ Literature Citations	25

Summary

The HuProt™ human proteome microarray provides the largest number of unique human proteins known to be included on a single slide, allowing thousands of interactions to be profiled in high-throughput.

Microarray type	Functional protein microarray
Species	Human
Production technology	Non-contact piezoelectric inkjet
Slide type	Coated glass*
Detection method	Fluorescence
Content	Recombinant human & mouse proteins**

*The HuProt™ human proteome microarray is available on two types of glass surfaces: PATH® and SuperEpoxy2™. Please contact us to discuss your needs.

** Please refer to the documentation accompanying the microarrays for specific information.

The HuProt™ version 3.1 microarray contains >19,500 unique proteins. This content encompasses 16,152 unique human genes (~81% of the proteome) and 124 unique mouse gene symbols. Recombinant proteins are expressed in the yeast *S. cerevisiae*, purified, and printed on glass slides in duplicate, along with control proteins (page 2).

These expressed recombinant proteins are N-terminal GST and RGS-His6-tagged, and the quality of each microarray batch is determined by GST immunoblotting (98% of all proteins show GST signals significantly higher than negative controls). For most applications, we print the microarray on glass slides that are coated with an ultra-thin layer of nitrocellulose film for the non-covalent, yet irreversible, capture of active proteins to the surface. The diameter of each HuProt spot (100 pL) is ~100 µm.

Summary, continued on page 2.



User Guide

HuProt™
Human Proteome Microarray v3.1

Page 2

TABLE OF CONTENTS

SUMMARY	1
Storage and Handling	3
Overview - Key Steps for HuProt™ Microarray Use	4
Additional Reagents and Materials	6
<hr/>	
ASSAYS	
Antibody Specificity	8
Antibody Specificity (Using Denatured Arrays)	11
Serum Profiling	14
RNA Binding	17
DNA Binding	19
Protein Binding	21
<hr/>	
APPENDIX	
Recipes	24
HuProt™ Literature Citations	25



Summary, continued from page 1.

HuProt™ V3.1 Control Spots

Control	Function
H1 - Histone H1 H2 (A+B) - Histone H2A and H2B mixture H3 - Histone H3 H4 - Histone H4	The histones are non-specific binding proteins used as positive controls for various assays, including serum profiling, antibody specificity assay, protein, DNA, RNA-binding assays, etc.
IgG488/594	Alexa Fluor 488/594 labeled IgG, positive control and landmarks for fluorescent detection in 488/594 channels.
Rhodamine + IgG 647	Rhodamine + Alexa Fluor 647 labeled IgG, positive control and landmarks for fluorescent detection in 532/635 channels.
Anti-human IgA	A positive control for human serum/plasma IgA profiling.
Mouse-anti-biotin	Detects biotin labeled protein probes and serves as a control for anti-mouse antibody detection reagent.
Rabbit-anti-biotin	Detects biotin labeled protein probes and serves as a control for anti-rabbit antibody detection reagent.
BSA - Bovine serum albumin	A negative control for non-specific protein interactions
Biotin-BSA - biotinylated BSA	A positive control for interaction with streptavidin-labeled detection reagent
Mouse IgM	Positive control for anti-mouse IgM detection.
RanBP2deltaFG	E3 SUMO-Protein Ligase, a positive control for the SUMOylation assay
hMDM2	E3 Ubiquitin Protein Ligase, a positive control for the ubiquitinylation assay
ERα	Estrogen receptor alpha. A positive control for ligand-binding assays.
Human IgM	A positive control for human serum/plasma IgM profiling.
Human IgG 1.5625 ng/μl Human IgG 6.25 ng/μl Human IgG 25 ng/μl Human IgG 100 ng/μl	The human IgG gradient serves as a positive control for human serum/plasma IgG profiling, and is used for Robust-Linear-Model Normalization in data analysis.
Anti-Human IgG 1.5625 ng/μl Anti-Human IgG 6.25 ng/μl Anti-Human IgG 25 ng/μl Anti-Human IgG 100 ng/μl	The Anti-human IgG gradient serves as a positive control for human serum/plasma IgG profiling, and is used for Robust-Linear-Model Normalization in data analysis.
GST (glutathione S-transferase) 10 ng/μl GST (glutathione S-transferase) 50 ng/μl GST (glutathione S-transferase) 100 ng/μl GST (glutathione S-transferase) 200 ng/μl	The glutathione S-transferase (GST) protein gradient serves as a negative control, and is used for background and statistical significance calculations.
Buffer	Printing buffer only, negative control

User Guide

HuProt™
Human Proteome Microarray v1.0

Page 3

TABLE OF CONTENTS

SUMMARY	1
Storage and Handling	3
Overview - Key Steps for HuProt™ Microarray Use	4
Additional Reagents and Materials	6
<hr/>	
ASSAYS	
Antibody Specificity	8
Antibody Specificity (Using Denatured Arrays)	11
Serum Profiling	14
RNA Binding	17
DNA Binding	19
Protein Binding	21
<hr/>	
APPENDIX	
Recipes	24
HuProt™ Literature Citations	26

Storage and Handling

IMPORTANT: New HuProt™ microarrays must be stored in an ultra cold and dry environment.

HuProt™ microarrays are shipped in closed plastic slide holders on dry ice, or with gel coolant sheets. Upon arrival, microarrays should immediately be stored at -80°C.

To ensure the best performance from the HuProt™ microarray:

- Wear gloves at all times
- Do not touch the active surface of the microarray (the surface where the bar code label is attached) with hands, with pipette tips or with tweezers. The active surface should face up at all times.
- Handle microarrays only along the edge near the barcode, using tweezers.
- When conducting assays, do not let the HuProt™ microarray dry out at any time.

User Guide

HuProt™
Human Proteome Microarray v3.1

Page 4

TABLE OF CONTENTS

SUMMARY	1
Storage and Handling	3
Overview - Key Steps for HuProt™ Microarray Use	4
Additional Reagents and Materials	6
<hr/>	
ASSAYS	
Antibody Specificity	8
Antibody Specificity (Using Denatured Arrays)	11
Serum Profiling	14
RNA Binding	17
DNA Binding	19
Protein Binding	21
<hr/>	
APPENDIX	
Recipes	24
HuProt™ Literature Citations	25



Overview - Key Steps for HuProt™ Microarray Use

Prepare a cold environment

- It is critical to keep the microarrays ultra-cold and ultra-dry right until they are used. HuProt™ microarrays should be stored at -80°C. Before removing HuProt™ microarrays from storage, place a layer of dry ice pellets inside a styrofoam box with a lid and cover tightly. Remove the plastic slide holders containing HuProt™ microarrays from the freezer and place it lengthwise on top of the dry ice. Add an additional layer of dry ice pellets or a sheet of frozen gel coolant on top of the plastic slide holders. Do not let condensate form on the surface before using the microarrays.

Prepare 4-well plates for the blocking step

- 4-well plates are used to hold the HuProt™ microarrays during blocking, during the assay (high volume samples) and during the blocking steps, the assay and during washing (e.g. *Nunc* 4-well rectangular dishes, Fisher Scientific No.12-565-495). Each compartment of the 4-well plates holds one microarray. To minimize cross-contamination of samples during the assay, some users prefer to use no more than 2 microarrays in each 4-well plate, keeping an empty well in between the microarrays.
- Add 3.0 ml of blocking buffer to each compartment of the 4-well plates (e.g. *Nunc* 4-well rectangular dishes, Fisher Scientific No.12-565-495).

IMPORTANT: Handle the microarray only on the edges of the glass, at the end where the bar code is attached. Do not touch the active surface of the microarray.

Block the HuProt™ microarrays

- The active surface of the HuProt™ microarray is the surface with the barcode. Carefully use fine-nosed tweezers to remove one microarray from the plastic slide holder that is resting on dry ice. Immediately submerge the HuProt™ microarray, with the active surface facing up, in a 4-well plate containing 3.0 of blocking buffer. Incubate with gentle shaking for 5 min.
- Carefully remove the blocking buffer from the 4-well plate (e.g. pour off the buffer, or use aspiration to remove it from a corner of the well). Use the microarrays immediately in the assay of choice (see below).

IMPORTANT: Blocking solution recipes vary by protocol—please refer to the relevant protocol for the correct recipe.

User Guide

HuProt™
Human Proteome Microarray v3.1

Page 5

TABLE OF CONTENTS

SUMMARY	1
Storage and Handling	3
Overview - Key Steps for HuProt™ Microarray Use	4
Additional Reagents and Materials	6
<hr/>	
ASSAYS	
Antibody Specificity	8
Antibody Specificity (Using Denatured Arrays)	11
Serum Profiling	14
RNA Binding	17
DNA Binding	19
Protein Binding	21
<hr/>	
APPENDIX	
Recipes	24
HuProt™ Literature Citations	25

Sample Preparation and Assay

- Dilute the sample to be tested in 3.0 ml of buffer.
- Carefully remove the blocking buffer from the 4-well plate.
- Immediately add the 3.0 ml of diluted sample to a compartment of the 4-well plate that contains a blocked HuProt™ microarray. The active surface of the microarray, with the barcode, should face up in the solution.

Washing

- Washes are carried out in the same 4-well plates in which the blocking and assay are conducted. Carefully remove the diluted sample from the 4-well plates (e.g. pour off the buffer, or use aspiration to remove it from a corner of the well). Add 4.0 ml of wash buffer to each compartment of the 4-well plate that contains a microarray (the buffer will vary depending on the assay).

IMPORTANT: If fluorescent probes are used, cover the plates with aluminum foil or use a lightproof storage container to minimize light exposure, which could quench the fluorescence of the labelled sample.

Scan and Store the Microarrays

- After the washes are completed, scan the microarrays immediately (highly preferred) or store them in a lightproof microscope slide box at -20°C. During storage, the active surface of the microarray should not touch any other surfaces, including other microarrays.

IMPORTANT: Microarrays must be scanned within 3 days after the assay is performed.



User Guide

HuProt™
Human Proteome Microarray v3.1

Page 6

TABLE OF CONTENTS

SUMMARY	1
Storage and Handling	3
Overview - Key Steps for HuProt™ Microarray Use	4
Additional Reagents and Materials	6
<hr/>	
ASSAYS	
Antibody Specificity	8
Antibody Specificity (Using Denatured Arrays)	11
Serum Profiling	14
RNA Binding	17
DNA Binding	19
Protein Binding	21
<hr/>	
APPENDIX	
Recipes	24
HuProt™ Literature Citations	25



Additional Reagents and Materials

Materials

Secondary Antibodies for Detection Step:

- The quality of commercially available secondary antibodies can vary widely. Please test all secondary antibodies thoroughly before conducting your assay (the secondary antibodies listed in this manual are provided as examples).
- Create grid alignment for data analysis using the GST-tagged proteins printed on the HuProt™ microarray as a visual reference.

IMPORTANT: We do not recommend doing the primary assay and anti-GST staining on the microarray at the same time. First conduct the primary assay, and scan the microarray data. After this, anti-GST probing can be done using an anti-Glutathione-S-Transferase antibody (e.g. Glutathione-S-Transferase, *S. japonicum* form; MED Millipore, cat no. AB3282). For this step, use a secondary antibody with a detection wavelength that is different from the wavelength used to detect the primary reaction.

NOTE: The GST signals obtained after your primary assay should not be used as a measure of the HuProt™ microarray quality, as all proteins on the microarray will become inactive after the primary assay. If too much anti-Glutathione-S-Transferase antibody is added, this may result in a very high background and may negatively affect data analysis.

Additional materials and equipment for incubations/assays:

- Aluminum foil
- Automatic pipettes
- Cleanroom wipes (preferred) or paper towels
- Micropipettes
- Orbital shaker
- Sterile disposable micropipette tips
- Sterile serological pipettes
- Fine nosed tweezers
- Vacuum system
- Vortex
- Plastic 4-well plates to store HuProt™ microarrays during the blocking, reaction and washing steps (e.g. Thermo Scientific *Nunc* Dishes, Rectangular 4-Well, No.12-565-495)
- BD Shoulder Cover Microscope Slide Box (lightproof; e.g. Fisher Scientific cat. no. 22-167-403)

User Guide

HuProt™
Human Proteome Microarray v3.1

Page 7

TABLE OF CONTENTS

SUMMARY	1
Storage and Handling	3
Overview - Key Steps for HuProt™ Microarray Use	4
Additional Reagents and Materials	6
<hr/>	
ASSAYS	
Antibody Specificity	8
Antibody Specificity (Using Denatured Arrays)	11
Serum Profiling	14
RNA Binding	17
DNA Binding	19
Protein Binding	21
<hr/>	
APPENDIX	
Recipes	24
HuProt™ Literature Citations	25

Additional Materials and Equipment for Data Analysis:

Microarray scanner (e.g. Molecular Devices GenePix 4000B) and computer
Microarray analysis program (e.g. GenePix Pro 6.1)

General Reagents:

Bovine Serum Albumin (IgG-Free, Protease-Free; Jackson ImmunoResearch Laboratories)

KCl

NaCl

Tris base

Tween-20

DTT (dithiothreitol) NOTE: DTT is not stable in solution. Only freshly-made DTT solutions should be used in buffers.

Triton-X 100

HEPES

Urea



User Guide

HuProt™
Human Proteome Microarray v3.1

Page 8

TABLE OF CONTENTS

SUMMARY	1
Storage and Handling	3
Overview - Key Steps for HuProt™ Microarray Use	4
Additional Reagents and Materials	6
<hr/>	
ASSAYS	
Antibody Specificity	8
Antibody Specificity (Using Denatured Arrays)	11
Serum Profiling	14
RNA Binding	17
DNA Binding	19
Protein Binding	21
<hr/>	
APPENDIX	
Recipes	24
HuProt™ Literature Citations	25

Assays Using the HuProt™ Human Proteome Microarray

I. Monoclonal Antibody Specificity Determination Assay

I.1 Storage:

- Store HuProt™ microarrays inside closed plastic slide holders at -80°C or on a layer of dry ice, right until the blocking step. The active surface of the HuProt™ microarray is the surface with the barcode.

IMPORTANT: It is critical to keep the HuProt™ microarrays ultra-cold and ultra-dry before use. Do not let condensate form on the surface before the microarray is used.

I.2 Blocking:

- Add 3.0 ml of blocking solution (5% BSA in 1X TBS-T) to each compartment of the 4-well plates. Carefully use fine-nosed tweezers to remove one microarray from the plastic slide holder that is resting on dry ice. Immediately submerge the HuProt™ microarray, active surface up, in a compartment of the 4-well plate containing blocking buffer. Incubate at room temperature, with gentle shaking for 1.5-2 hrs.

I.3 Sample Preparation: Prepare the primary monoclonal antibody for testing

- Dilute the primary antibody to a concentration of 1 µg/ml in 3.0 ml blocking buffer. Store on ice.
- If the antibody of interest is in supernatant, dilute the supernatant 1:12 in blocking buffer to a final volume of 3.0 ml. Store on ice.

NOTE: This dilution assumes that the antibody concentration in the supernatant is 5 µg /ml. Please dilute based on the actual antibody concentration in your sample.



User Guide

HuProt™
Human Proteome Microarray v3.1

Page 9

TABLE OF CONTENTS

SUMMARY	1
Storage and Handling	3
Overview - Key Steps for HuProt™ Microarray Use	4
Additional Reagents and Materials	6
<hr/>	
ASSAYS	
Antibody Specificity	8
Antibody Specificity (Using Denatured Arrays)	11
Serum Profiling	14
RNA Binding	17
DNA Binding	19
Protein Binding	21
<hr/>	
APPENDIX	
Recipes	24
HuProt™ Literature Citations	25

- I.4** Assay: Add monoclonal antibody sample to HuProt™ microarrays
- Carefully remove the blocking buffer from the 4-well plates that contain the HuProt™ microarrays. Do not touch the microarray surface.
 - Carefully add 3.0 ml of diluted sample to the active (top) surface of the microarray in the 4-well plate. Be careful not to touch the surface of the microarray. Incubate with gentle shaking on an orbital shaker at room temperature for 1 hr.
- I.5** Washing:
- When the assay is completed, carefully remove the assay buffer from the 4-well plates that contain the HuProt™ microarrays.
 - Rinse each slide briefly with 4.0 ml of 1X TBS-T. Carefully remove the buffer after rinsing.
 - Add 4.0 ml of 1X TBS-T and incubate with gentle shaking at room temperature for 10 min, then carefully remove the buffer. Repeat for a total of three washes.
- I.6** Detection: Add Secondary Antibody
- Dilute secondary/detection antibody in blocking buffer to the manufacturer-recommended concentration.
 - Add 3.0 ml of diluted secondary antibody to each compartment in a 4-well plate containing the HuProt™ microarrays.
 - Cover the 4-well plates with aluminum foil and incubate at room temperature for 1 hr with gentle shaking.
 - Carefully remove the buffer containing the secondary antibody from a corner of the 4-well plate by aspiration.

IMPORTANT: After adding the secondary antibody, store the microarrays in the dark. For all incubation and washing steps below, cover the 4-well plates containing HuProt™ microarrays with aluminum foil to minimize light exposure, which can quench the fluorescence.



User Guide

HuProt™
Human Proteome Microarray v3.1

Page 10

TABLE OF CONTENTS

SUMMARY	1
Storage and Handling	3
Overview - Key Steps for HuProt™ Microarray Use	4
Additional Reagents and Materials	6

ASSAYS

Antibody Specificity	8
Antibody Specificity (Using Denatured Arrays)	11
Serum Profiling	14
RNA Binding	17
DNA Binding	19
Protein Binding	21

APPENDIX

Recipes	24
HuProt™ Literature Citations	25

Reference:

Jeong JS *et al.* (2012) Rapid identification of monospecific monoclonal antibodies using a human proteome microarray. *Mol Cell Proteomics* 11 10.1074/mcp.O111.016253.

Liu S *et al.* (2015) Characterization of monoclonal antibody's binding kinetics using oblique-incidence reflectivity difference approach. *MAbs* 7(1):110-9. doi: 10.4161/19420862.2014.985919.

I.7 Washing:

- Carefully remove the buffer containing the secondary antibody. Rinse each slide briefly with 4.0 ml of 1X TBS-T. Carefully remove the buffer after rinsing.
- Add 4.0 ml of 1X TBS-T to each compartment of the 4-well plate, cover with foil, and incubate with gentle shaking for 10 min at room temperature.
- Briefly rinse the microarrays three times with 0.1X TBS.

I.8 Drying:

- Place clean room wipes or paper towels on the bottom of a black microscope slide box (will hold several microarrays) or prepare plastic conical 50-ml tubes (each will hold one microarray).
- Remove the HuProt™ microarray from the 4-well plates and tap the edge lightly on a clean room wipe to remove excess fluid. Do not touch the active surface of the microarray. Carefully slot the microarrays into the microscope slide box – the microarrays will be perpendicular to the paper towels lining the base of the box. If you are using 50-ml tubes, carefully slide a single microarray lengthwise into the tube.
- To remove excess fluid, spin the microscope slide box or the 50-ml tubes containing the microarrays in a centrifuge at 800 rpm for 3 min (spinning at higher speeds may break the microarray). After centrifugation, carefully remove the HuProt™ microarrays and discard the clean room wipes.

I.9 Scanning:

- The HuProt™ microarrays can be scanned immediately (highly preferred) or stored at -20°C in a lightproof box.

IMPORTANT: Microarrays must be scanned within 3 days after the assay is performed.

Appendix for Monoclonal Antibody Specificity Determination:

Recipes

- 1X TBS-T Buffer: TBS pH 7.5, 0.1% Tween-20 (see General Appendix, Buffers section)
- Blocking Solution: 5% BSA [w/v] in 1X TBS-T Buffer: Dissolve 5 g of Bovine Serum Albumin (IgG-Free, Protease-Free) in 60 ml of 1X TBS-T. Complete to 100 ml with 1X TBS-T and filter before use.

User Guide

HuProt™
Human Proteome Microarray v3.1

Page 11

TABLE OF CONTENTS

SUMMARY	1
Storage and Handling	3
Overview - Key Steps for HuProt™ Microarray Use	4
Additional Reagents and Materials	6
<hr/>	
ASSAYS	
Antibody Specificity	8
Antibody Specificity (Using Denatured Arrays)	11
Serum Profiling	14
RNA Binding	17
DNA Binding	19
Protein Binding	21
<hr/>	
APPENDIX	
Recipes	24
HuProt™ Literature Citations	25



II. Monoclonal Antibody Specificity Determination Assay Using Denatured Arrays

II.1 Storage:

- Store HuProt™ microarrays inside closed plastic slide holders at -80°C or on a layer of dry ice, right until the blocking step. The active surface of the HuProt™ microarray is the surface with the barcode.

IMPORTANT: It is critical to keep the HuProt™ microarrays ultra-dry before use. Do not let condensate form on the surface before the microarray is used.

II.2 Blocking:

- Add 3.0 ml of blocking solution (5% BSA in 1X TBS-T) to each compartment of the 4-well plates. Carefully use fine-nosed tweezers to remove one microarray from the plastic slide holder that is resting on dry ice. Immediately submerge the HuProt™ microarray, active surface up, in a compartment of the 4-well plate containing blocking buffer. Incubate at room temperature, with gentle shaking for 1.5-2 hrs.

II.3 Denature the HuProt™ Array:

- Add 3.0 ml of denaturation buffer (9M Urea with 5 mM DTT) to each compartment of the 4-well plates. Incubate at room temperature for 20 min WITHOUT shaking.
- Wash in 1X TBS-T for 5 min with gentle shaking. Carefully remove the wash buffer. Add 3.0 ml of fresh 1X TBS-T and repeat, for a total of 3 washes.

II.4 Sample Preparation: Prepare the primary monoclonal antibody for testing

- Dilute the primary antibody to a concentration of 1 µg/ml in 3.0 ml blocking buffer. Store on ice.
- If the antibody of interest is in supernatant, dilute the supernatant 1:12 in blocking buffer to a final volume of 3.0 ml. Store on ice.

NOTE: This dilution assumes that the concentration of antibody in the supernatant is 5 µg/ml. Please dilute based on the actual antibody concentration in your sample.

II.5 Assay: Add monoclonal antibody sample to HuProt™ microarrays

- Carefully remove the blocking buffer from the 4-well plates that contain the blocked HuProt™ microarrays. Do not touch the microarray surface.
- Carefully add 3.0 ml sample to the active (top) surface of the microarray in the 4-well plate. Be careful not to touch the surface of the microarray. Incubate with gentle shaking on an orbital shaker at room temperature for 1 hr, or at 4°C overnight.

User Guide

HuProt™
Human Proteome Microarray v3.1

Page 12

TABLE OF CONTENTS

SUMMARY	1
Storage and Handling	3
Overview - Key Steps for HuProt™ Microarray Use	4
Additional Reagents and Materials	6
<hr/>	
ASSAYS	
Antibody Specificity	8
Antibody Specificity (Using Denatured Arrays)	11
Serum Profiling	14
RNA Binding	17
DNA Binding	19
Protein Binding	21
<hr/>	
APPENDIX	
Recipes	24
HuProt™ Literature Citations	25



II.6 Washing:

- Upon completing the assay, carefully remove the assay buffer from the 4-well plates that contain the HuProt™ microarrays.
- Rinse each slide briefly with 4.0 ml of 1X TBS-T. Carefully remove the buffer.
- Add 4.0 ml of 1X TBS-T and incubate with gentle shaking at room temperature for 10 min, then carefully remove the buffer. Repeat for a total of three washes.

II.7 Detection: Add Secondary Antibody

- Dilute secondary/detection antibody in blocking buffer to the manufacturer-recommended concentration.
- Add 3.0 ml of diluted secondary antibody to each compartment of a 4-well plate that contains HuProt™ microarrays.
- Cover the 4-well plates with aluminum foil and incubate at room temperature for 1 hr with gentle shaking.

IMPORTANT: After the secondary antibody is added, store the microarrays in the dark. For all incubation and washing steps below, cover the 4-well plates containing HuProt™ microarrays with aluminum foil to minimize light exposure, which can quench the fluorescence.

II.8 Washing:

- Carefully remove the buffer containing the secondary antibody.
- Rinse each slide briefly with 4.0 ml of 1X TBS-T. After rinsing, carefully remove the buffer.
- Add 4.0 ml 1X TBS-T to each compartment of the 4-well plate, cover with foil, and incubate with gentle shaking for 10 min at room temperature. Carefully remove the buffer. Repeat for a total of three washes.
- Briefly rinse the microarrays three times with 0.1X TBS

II.9 Drying:

- Place clean room wipes or paper towels on the bottom of a black microscope slide box (will hold several microarrays) or prepare plastic conical 50-ml tubes (each will hold one microarray).
- Remove the HuProt™ microarray from the 4-well plates and tap the edge lightly on a clean room wipe to remove excess fluid. Do not touch the active surface of the microarray. Carefully slot the microarrays into the microscope slide box – the microarrays will be perpendicular to the paper towels lining the base of the box. If you are using 50 ml tubes, carefully slide a single microarray lengthwise into the tube.

User Guide

HuProt™
Human Proteome Microarray v3.1

Page 13

TABLE OF CONTENTS

SUMMARY	1
Storage and Handling	3
Overview - Key Steps for HuProt™ Microarray Use	4
Additional Reagents and Materials	6
<hr/>	
ASSAYS	
Antibody Specificity	8
Antibody Specificity (Using Denatured Arrays)	11
Serum Profiling	14
RNA Binding	17
DNA Binding	19
Protein Binding	21
<hr/>	
APPENDIX	
Recipes	24
HuProt™ Literature Citations	25

- To remove excess fluid, spin the microscope slide box or the 50 ml tubes containing the microarrays in a centrifuge at 800 rpm for 3 min (spinning at higher speeds may break the microarray). After centrifugation, carefully remove the HuProt™ microarrays and discard the clean room wipes.

II.10 Scanning:

- The HuProt™ microarrays can be scanned immediately (highly preferred) or stored at -20°C in a lightproof box.

IMPORTANT: Microarrays must be scanned within 3 days after the assay is performed.

Appendix for Monoclonal Antibody Specificity Determination:

Recipes

- Denaturation Buffer: 9M Urea with 5 mM DTT in water. Dissolve urea in water and filter; add freshly-made DTT to the solution immediately before use.
- 1X TBS-T Buffer: 1X TBS pH 7.5, 0.1% Tween-20 (see General Appendix, Buffers section)
- Blocking Solution: 5% BSA [w/v] in 1X TBS-T Buffer: Dissolve 5 g of Bovine Serum Albumin (IgG-Free, Protease-Free) in 60 ml of 1X TBS-T. Complete to 100 ml with 1X TBS-T and filter before use.



User Guide

HuProt™
Human Proteome Microarray v3.1

Page 14

TABLE OF CONTENTS

SUMMARY	1
Storage and Handling	3
Overview - Key Steps for HuProt™ Microarray Use	4
Additional Reagents and Materials	6
<hr/>	
ASSAYS	
Antibody Specificity	8
Antibody Specificity (Using Denatured Arrays)	11
Serum Profiling	14
RNA Binding	17
DNA Binding	19
Protein Binding	23
<hr/>	
APPENDIX	
Recipes	24
HuProt™ Literature Citations	25



Assays Using the HuProt™ Human Proteome Microarray

III. Serum Profiling Assay

SERUM SAMPLE STORAGE: Repeated thawing and re-freezing of serum samples can affect the reproducibility of serum profiling data obtained from HuProt™ arrays. When collecting serum samples, divide the serum into single-use aliquots and store them at -80°C. This will optimize the reproducibility of the experiments.

NOTE: Typically a 1:1000 dilution of serum is used.

EPITOPE SPREADING: Clients who have studied epitope spreading on HuProt™ arrays recommend using a 1:500 dilution of serum.

III.1 Storage:

- Store HuProt™ microarrays in closed plastic slide holders at -80°C or on a layer of dry ice, right until the blocking step. The active surface of the HuProt™ microarray is the surface with the barcode.

IMPORTANT: It is critical to keep the HuProt™ microarrays ultra cold and ultra dry prior to use. Do not let liquid condense onto the microarray surface before use.

III.2 Blocking:

- Add 3.0 ml of blocking solution (5% BSA/1X TBS-T) to each compartment in the 4-well plates.
- Carefully use fine-nosed tweezers to remove one microarray from the plastic slide holder that has been resting on dry ice. Immediately submerge the HuProt™ microarray, active side up, in a 4-well plate containing blocking buffer. Incubate with gentle shaking for 1.5-2 hrs at room temperature.

III.3 Sample preparation:

- The recommended serum dilution is 1:1000 in blocking buffer.
- Dilute the primary sample to a final volume of 3.0 ml in blocking buffer, then vortex briefly and store on ice.

III.4 Assay: Add Sample to HuProt™ Microarray

- Carefully remove the blocking buffer from the 4-well plates containing the HuProt™ microarrays.

User Guide

HuProt™
Human Proteome Microarray v3.1

Page 15

TABLE OF CONTENTS

SUMMARY	1
Storage and Handling	3
Overview - Key Steps for HuProt™ Microarray Use	4
Additional Reagents and Materials	6
<hr/>	
ASSAYS	
Antibody Specificity	8
Antibody Specificity (Using Denatured Arrays)	11
Serum Profiling	14
RNA Binding	17
DNA Binding	19
Protein Binding	21
<hr/>	
APPENDIX	
Recipes	24
HuProt™ Literature Citations	25



- Carefully pipette the prepared serum sample (3.0 ml) onto the active surface of the blocked HuProt™ microarray (barcode facing up). Be careful not to let the pipette tip touch the microarray surface.
- Incubate with gentle shaking on an orbital shaker overnight at 4°C (e.g. in cold room).

III.5 Washing:

- Following incubation, carefully remove the diluted serum from the 4-well plate. Add 4.0 ml of 1X TBS-T to each well and rinse briefly. After rinsing, carefully remove the buffer.
- Add 4.0 ml of 1X TBS-T and incubate 10 min at room temperature with gentle shaking. Remove the buffer by aspiration from the corner of the 4-well plate. Repeat for a total of three 10 min washes. Do not let the HuProt™ microarray dry out at any time.

IMPORTANT: For all steps below, cover the 4-well plates containing the HuProt™ microarrays tightly with aluminum foil during all incubations and washes to minimize light exposure.

III.6 Detection: Add Secondary Antibody

- Dilute the secondary antibody in blocking buffer per manufacturer’s directions for Western Blot use. Add 3.0 ml of the diluted secondary antibody to each freshly washed HuProt™ microarray. Cover the 4-well plate with aluminum foil and incubate for 1-1.5 hrs with gentle shaking at room temperature.

III.7 Washing:

- Carefully remove the buffer containing the secondary antibody. Add 4.0 ml 1X TBS-T to briefly rinse the microarray. Carefully remove the rinse buffer.
- Add 4.0 ml of 1X TBS-T and incubate for 10 min at room temperature with gentle shaking. Carefully remove the wash buffer from the 4-well plate. Repeat for a total of three 10 min washes.
- Briefly rinse the microarrays three times with 0.1X TBS.

IMPORTANT: Do not let the HuProt™ microarrays sit for long periods in wash buffer. Proceed to the drying stage immediately.

Alternative Final Rinse method:

- Hold the array on the barcoded edge with tweezers. Dip the array rapidly in a fresh beaker of ddH₂O three times. Gently blot the edge of the washed array on kimwipes or clean room wipes, and then proceed to the drying step.

IMPORTANT: Proceed to the drying stage immediately.

User Guide

HuProt™
Human Proteome Microarray v3.1

Page 16

TABLE OF CONTENTS

SUMMARY	1
Storage and Handling	3
Overview - Key Steps for HuProt™ Microarray Use	4
Additional Reagents and Materials	6
<hr/>	
ASSAYS	
Antibody Specificity	8
Antibody Specificity (Using Denatured Arrays)	11
Serum Profiling	14
RNA Binding	17
DNA Binding	19
Protein Binding	21
<hr/>	
APPENDIX	
Recipes	24
HuProt™ Literature Citations	25

II.8 Drying:

- Place clean room wipes or paper towels on the bottom of a black microscope slide box (will hold several microarrays) or prepare plastic conical 50-ml tubes (each will hold one microarray).
- Remove the HuProt™ microarray from the 4-well plates and tap the edge lightly on a clean room wipe or paper towel to remove excess fluid. Do not touch the active surface of the microarray. Carefully slot the microarrays into the microscope slide box – the microarrays will be perpendicular to the paper towels lining the box. If you are using 50-ml tubes, carefully slide a single microarray lengthwise into the tube.
- Spin the microscope slide box or the 50-ml tubes containing the microarrays in a centrifuge at 800 rpm for 3 min to remove excess fluid (spinning at higher speeds may break the microarrays). After centrifugation, carefully remove the HuProt™ microarrays and discard the clean room wipe or paper towel.

III.9 Scanning:

- The HuProt™ microarray can be scanned immediately (highly preferred), or stored at -20°C in a lightproof box.

IMPORTANT: Microarrays must be scanned within 3 days after the assay is performed.

Appendix for Serum Profiling Assay:

Recipes

- 1X TBS-T Buffer: TBS pH 7.5, 0.1% Tween-20 (see General Appendix, buffer recipe)
- Blocking Solution: 5% BSA [w/v] in 1X TBS-T

References:

- Hu CJ *et al.* (2012) Identification of new auto-antigens for primary biliary cirrhosis using human proteome microarrays. *Mol Cell Proteomics*, **11**(9): 669-80.
- Gupta S *et al.* (2017) Evaluation of autoantibody signatures in meningioma patients using human proteome arrays. *Oncotarget* 2017 Apr 10. doi: 10.18632/oncotarget.16997. [Epub ahead of print]
- Hu C *et al.* (2015) Autoantibody Profiling on Human Proteome Microarray for Biomarker Discovery in Cerebrospinal Fluid and Sera of Neuropsychiatric Lupus. *PLoS One* **10**(5): e0126643.
- Chung, BK *et al.* (2017) Phenotyping and auto-antibody production by liver-infiltrating B cells in primary sclerosing cholangitis and primary biliary cholangitis. *J Autoimmunity* **77**: 45-54.
- Syed P *et al.* (2015) Autoantibody Profiling of Glioma Serum Samples to Identify Biomarkers Using Human Proteome Arrays. *Sci Rep.* 2015 Sep **15**:5:13895. doi: 10.1038/srep13895.
- Shi L *et al.* (2016). Application of high-throughput protein array in clinical screening for tumor markers. *Int J Clin Exp Med* **9**(5):8529-8535.
- Yang L *et al.* (2016) Identification of serum biomarkers for gastric cancer diagnosis using a human proteome microarray. *Mol Cell Proteomics* **15**(2):614-23. doi: 10.1074/mcp.M115.051250. Epub 2015 Nov 23
- Iyama S *et al.* (2016) Drebrin: A new oncofetal biomarker associated with prognosis of lung adenocarcinoma. *Lung Cancer* **102**:74-81; doi: 10.1016/j.lungcan.2016.10.013.

User Guide

HuProt™
Human Proteome Microarray v3.1

Page 17

TABLE OF CONTENTS

SUMMARY	1
Storage and Handling	3
Overview - Key Steps for HuProt™ Microarray Use	4
Additional Reagents and Materials	6
<hr/>	
ASSAYS	
Antibody Specificity	8
Antibody Specificity (Using Denatured Arrays)	11
Serum Profiling	14
RNA Binding	17
DNA Binding	19
Protein Binding	21
<hr/>	
APPENDIX	
Recipes	24
HuProt™ Literature Citations	25



Assays Using the HuProt™ Human Proteome Microarray

IV. Fluorescently Labeled RNA Binding Assay

IV.1 Storage:

- Store HuProt™ microarrays in plastic slide holders at -80°C or on a layer of dry ice, right until the blocking step. The active surface of the HuProt™ microarray is the surface with the barcode.

IMPORTANT: It is critical to keep the HuProt™ microarrays ultra-cold and ultra-dry before use. Do not let condensate form on the surface before the microarray is used.

IV.2 Blocking:

- Add 3.0 ml of blocking solution (recipe on Pg. 19) to each compartment of the 4-well plates. Carefully use fine-nosed tweezers to remove one microarray from the plastic slide holder that is resting on dry ice. Immediately submerge the HuProt™ microarray, active surface up, in a compartment of the 4-well plate containing blocking buffer. Incubate at room temperature, with gentle shaking for 1.5-2 hrs.

IV.3 Sample preparation:

- Dilute the fluorescently labeled RNA sample to 250 nM in 3.0 ml RNA binding buffer.

IV.4 Assay: Add Labeled RNA to HuProt™ microarrays

- Carefully remove the blocking buffer from the 4-well plates.
- Carefully pipette 3.0 ml of diluted RNA sample onto the active surface of the blocked HuProt™ microarray. Be careful not to let the pipette tip touch the active surface.

IV.5 Washes:

- Carefully remove the assay buffer from the 4-well plates.
- Add 4.0 ml 1X TBS-T to each compartment in the 4-well plates. Shake gently for 5 min at room temperature, and then carefully remove the wash buffer. Repeat for a total of four washes.

IMPORTANT: Cover the 4-well plates with aluminum foil at all times to minimize light exposure.

- Briefly rinse the microarrays three times with 0.1X TBS.
- Remove the HuProt™ microarrays from the 4-well plates with tweezers. Tap the edge of the microarray lightly on a clean room wipe to remove excess fluid. Do not touch the active surface of the microarray.

User Guide

HuProt™
Human Proteome Microarray v3.1

Page 18

TABLE OF CONTENTS

SUMMARY	1
Storage and Handling	3
Overview - Key Steps for HuProt™ Microarray Use	4
Additional Reagents and Materials	6
<hr/>	
ASSAYS	
Antibody Specificity	8
Antibody Specificity (Using Denatured Arrays)	11
Serum Profiling	14
RNA Binding	17
DNA Binding	19
Protein Binding	21
<hr/>	
APPENDIX	
Recipes	24
HuProt™ Literature Citations	25



IV.6 Drying:

- Place clean room wipes or paper towels on the bottom of a black microscope slide box (will hold several microarrays) or prepare plastic conical 50 ml tubes (each will hold one microarray).
- Remove the HuProt™ microarray from the 4-well plates and tap the edge lightly on a clean room wipe to remove excess fluid. Do not touch the active surface of the microarray. Carefully slot the microarrays into the microscope slide box – the microarrays will be perpendicular to the wipes lining the box. If you are using 50 ml tubes, carefully slide a single microarray lengthwise into the tube.
- To remove excess fluid, spin the microscope slide box or the 50 ml tubes containing the microarrays in a centrifuge at 800 rpm for 3 min (spinning at higher speeds may break the microarray). After centrifugation, carefully remove the HuProt™ microarrays and discard the clean room wipes.

IV.7 Scanning and Storage:

- The HuProt™ microarrays can be scanned immediately (highly preferred), or stored at -20°C in a lightproof box.

IMPORTANT: Microarrays must be scanned within 3 days after the assay

Appendix for Fluorescently-labeled RNA-binding Assay

Recipes

- 1x TBS-T (See General Appendix)
- Blocking Solution: 10 µg/ml ssDNA (herring sperm); 2 mM MgCl₂; 5 mg/ml BSA in SuperBlock T20 (PBS) Blocking Buffer (Thermo Scientific Cat. No. 37516)
- RNA Binding Buffer: 2 mM MgCl₂; 2 mg/ml BSA; 10 µg/ml ssDNA (herring sperm) in 1X PBS

References:

Barry G *et al.* (2013). The long non-coding RNA Gomafu is acutely regulated in response to neuronal activation and involved in schizophrenia-associated alternative splicing. *Molec Psychiatry* **Apr 30**. doi: 10.1038/mp.2013.45. ePub ahead of print.

Donnelly CJ *et al.* (2013). RNA Toxicity from the ALS/FTD C9orf72 Expansion is Mitigated by Antisense Intervention. *Neuron* **80(2)**: 415-28.

Fan B *et al.* (2013). A human microarray identifies that the heterogeneous nuclear ribonucleoprotein K (hnRNP K) recognizes the 5' terminal sequence of the hepatitis C virus RNA. *Mol Cell Proteomics*. **13(1)**:84-92.

User Guide

HuProt™
Human Proteome Microarray v3.1

Page 19

TABLE OF CONTENTS

SUMMARY	1
Storage and Handling	3
Overview - Key Steps for HuProt™ Microarray Use	4
Additional Reagents and Materials	6
<hr/>	
ASSAYS	
Antibody Specificity	8
Antibody Specificity (Using Denatured Arrays)	11
Serum Profiling	14
RNA Binding	17
DNA Binding	19
Protein Binding	21
<hr/>	
APPENDIX	
Recipes	24
HuProt™ Literature Citations	25



Assays Using the HuProt™ Human Proteome Microarray

V. Fluorescently Labeled DNA Binding Assay

V.1 Storage:

- Store HuProt™ microarrays in closed plastic slide holders at -80°C or on a layer of dry ice, right until the blocking step. The active surface of the HuProt™ microarray is the surface with the barcode.

IMPORTANT: It is critical to keep the HuProt™ microarrays ultra-cold and ultra-dry before use. Do not let condensate form on the surface before the microarray is used.

V.2 Blocking:

- Add 3.0 ml of DNA hybridization buffer (pre-chilled on ice) to each compartment of the 4-well plates.
- Use fine-nosed tweezers to carefully remove one microarray from the plastic slide holder that is resting on dry ice. Immediately submerge the HuProt™ microarray, active surface up, into a well that contains blocking buffer.
- Incubate with gentle shaking for 3 hrs at 4°C.

V.3 Sample Preparation:

- Dilute the fluorescently labeled DNA sample to 40 nM in 3.0 ml of DNA hybridization buffer containing poly (dA-dT).

V.4 Assay: Add Labeled DNA to the HuProt™ microarrays

- After blocking, carefully remove the blocking buffer from the 4-well plate.
- Immediately add 3.0 ml of diluted DNA sample to a well containing a blocked HuProt™ microarray. Do not let the pipette tip touch the microarray surface.
- Cover the 4-well plate with aluminum foil to minimize exposure to light. Incubate at 4°C overnight with gentle shaking.

V.5 Washing:

- One wash cycle is sufficient for this assay.
- Carefully remove the assay buffer from the 4-well plate containing the microarrays.

User Guide

HuProt™
Human Proteome Microarray v3.1

Page 20

TABLE OF CONTENTS

SUMMARY	1
Storage and Handling	3
Overview - Key Steps for HuProt™ Microarray Use	4
Additional Reagents and Materials	6
<hr/>	
ASSAYS	
Antibody Specificity	8
Antibody Specificity (Using Denatured Arrays)	11
Serum Profiling	14
RNA Binding	17
DNA Binding	19
Protein Binding	21
<hr/>	
APPENDIX	
Recipes	24
HuProt™ Literature Citations	25



- Add 4.0 ml of pre-chilled wash buffer to the well containing the microarray. Incubate with gentle shaking for 1-3 min at 4°C.
- Use tweezers to remove the HuProt™ microarrays from the wash buffer and tap the edge lightly on a clean room wipe or paper towel to remove excess buffer. Do not touch the active surface of the microarray.

V.6 Drying:

- Place clean room wipes or paper towels on the bottom of a black microscope slide box (will hold several microarrays) or prepare plastic conical 50-ml tubes (each will hold one microarray).
- Remove the HuProt™ microarray from the 4-well plates and tap the edge lightly on a clean room wipe to remove excess fluid. Do not touch the active surface of the microarray. Carefully slot the microarrays into the microscope slide box, placing the microarrays perpendicular to the paper towels lining the base of the box. If you are using 50-ml tubes, carefully slide a single microarray lengthwise into the tube.
- Spin the microscope slide box or the 50-ml tubes containing the microarrays in a centrifuge at 800 rpm for 3 min to remove excess fluid (spinning at higher speeds may break the microarray). After centrifugation, carefully remove the HuProt™ microarrays and discard the clean room wipes.

V.7 Scanning and Storage:

- The HuProt™ microarrays can be scanned immediately (highly recommended), or stored for up to three days at -20°C in a lightproof box.

IMPORTANT: Microarrays must be scanned within 3 days after the assay is performed.

Appendix for Fluorescently Labeled DNA Binding Assay

Recipes

- DNA hybridization buffer (store on ice or at 4°C): 25 mM HEPES at pH 8.0; 50 mM potassium glutamate; 0.1% Triton X-100; 8 mM magnesium acetate; 3 mM DTT (freshly made and added to the buffer right before use); 4 μM poly (dA-dT); 10% glycerol
- Wash buffer: Same as above, but without 4 μM poly (dA-dT)

Reference:

Hu S *et al.* (2009) Profiling the human protein-DNA interactome reveals ERK2 as a transcriptional repressor of interferon signaling. *Cell* **139**(3): 610-22.

User Guide

HuProt™
Human Proteome Microarray v3.1

Page 21

TABLE OF CONTENTS

SUMMARY	1
Storage and Handling	3
Overview - Key Steps for HuProt™ Microarray Use	4
Additional Reagents and Materials	6
<hr/>	
ASSAYS	
Antibody Specificity	8
Antibody Specificity (Using Denatured Arrays)	11
Serum Profiling	14
RNA Binding	17
DNA Binding	19
Protein Binding	21
<hr/>	
APPENDIX	
Recipes	24
HuProt™ Literature Citations	25



Assays Using the HuProt™ Human Proteome Microarray

VI. Fluorescently Labeled Protein Binding Assay

VI.1 Storage:

- Store HuProt™ microarrays in closed plastic slide holders at -80°C or on a layer of dry ice, right until the blocking step. The active surface of the HuProt™ microarray is the surface with the barcode.

IMPORTANT: It is critical to keep the HuProt™ microarrays ultra-cold and ultra-dry before use. Do not let condensate form on the surface before the microarray is used.

VI.2 Blocking:

- Add 3.0 ml of blocking solution (5% BSA in 1X TBS-T) to each compartment of the 4-well plates. Carefully use fine-nosed tweezers to remove one microarray from the plastic slide holder that is resting on dry ice. Immediately submerge the HuProt™ microarray, active surface up, in a compartment of the 4-well plate containing blocking buffer. Incubate at room temperature with gentle shaking for 1.5-2 hrs.

VI.3 Sample Preparation:

- Dilute 3 µg fluorescently labeled sample protein in 3.0 ml blocking buffer.

VI.4 Assay: Add Labeled Protein to HuProt™ microarrays

- Carefully remove the blocking buffer from the 4 well plates that contain microarrays. Add 3.0 ml of diluted labeled protein (3 µg protein in 3.0 ml blocking solution) to each well that contains a blocked HuProt™ microarray (active side up).
- Cover the plates with aluminum foil to minimize light exposure, which could quench the fluorescence of the protein sample. Incubate with gentle shaking for 1 hr at room temperature.

VI.5 Washing:

- Carefully remove the assay buffer from the 4-well plates. Add 4.0 ml 1X TBS-T to briefly rinse the microarray. After rinsing, carefully remove the buffer.
- Add 4.0 ml 1X TBS-T buffer and wash with gentle shaking for 10 minutes. Carefully remove the buffer. Repeat for a total of three long washes.
- Briefly rinse the microarrays three times with 0.1X TBS.

User Guide

HuProt™
Human Proteome Microarray v3.1

Page 22

TABLE OF CONTENTS

SUMMARY	1
Storage and Handling	3
Overview - Key Steps for HuProt™ Microarray Use	4
Additional Reagents and Materials	6
<hr/>	
ASSAYS	
Antibody Specificity	8
Antibody Specificity (Using Denatured Arrays)	11
Serum Profiling	14
RNA Binding	17
DNA Binding	19
Protein Binding	21
<hr/>	
APPENDIX	
Recipes	24
HuProt™ Literature Citations	25



For research use only.
Not for use in diagnostic procedures.

- **IMPORTANT:** Cover the 4-well plates with aluminum foil at all times during both the reaction and washing steps to minimize light exposure, which could quench the fluorescence of the labeled protein sample.

VI.6 Drying:

- Place clean room wipes or paper towels on the bottom of a black microscope slide box (will hold several microarrays) or prepare plastic conical 50-ml tubes (each will hold one microarray).
- Remove the HuProt™ microarray from the 4-well plate and tap the edge lightly on a paper towel to remove excess fluid. Do not touch the active surface of the microarray. Carefully slot the microarrays into the microscope slide box - the microarrays will be perpendicular to the paper towels lining the box. If you are using conical plastic 50-ml tubes, carefully slide a single microarray lengthwise into the tube.
- To remove excess fluid, spin the microscope slide box or the 50-ml tubes containing the microarrays in a centrifuge at 800 rpm for 3 min (spinning at higher speeds may break the microarray). After centrifugation, carefully remove the HuProt™ microarrays and discard the clean room wipe.

VI.7 Scanning and Storage:

- The HuProt™ microarrays can be scanned immediately (highly preferred) or stored at -20°C in a lightproof box.

IMPORTANT: Microarrays must be scanned within 3 days after the assay is performed.

Appendix for protein binding assay using fluorescently-labeled protein probes

Recipes

- 1X TBS-T (see General Appendix)
- Blocking solution (5% BSA in 1X TBS-T)

References:

- Huang Y *et al.* (2012). Global tumor protein p53/p63 Interactome: making a case for cisplatin chemoresistance. *Cell Cycle* **11(12)**:2367-79.
- Chen Y. *et al.* (2013). Bcl2-associated Athanogene 4 Interactive Analysis Reveals a New Role in Modulating Proteosome Activity. *Mol Cell Proteomics* **12(10)**:2804-19.
- Fan, Q *et al.* (2014). Identification of proteins that interact with alpha A-crystallin using a human proteome microarray. *Mol Vis* **20**: 117–124. eCollection 2014.
- Deng, RP *et al.* (2014). Global Identification of O-GlcNAc Transferase (OGT) Interactors by a Human Proteome Microarray and the Construction of an OGT Interactome. *Proteomics* 2014 **14(9)**:1020-30. doi: 10.1002/pmic.201300144. Epub 2014 Mar 25.
- Ma TM *et al.* (2014) Serine Racemase Regulated by Binding to Stargazin and PSD-95: Potential NMDA-AMPA Glutamate Neurotransmission Cross-talk. *J Biol Chem* pii: **jbic.M114.571604**. [Epub ahead of print]; First Published on August 27, 2014.

Continued...

User Guide

HuProt™
Human Proteome Microarray v3.1

Page 23

TABLE OF CONTENTS

SUMMARY	1
Storage and Handling	3
Overview - Key Steps for HuProt™ Microarray Use	4
Additional Reagents and Materials	6
<hr/>	
ASSAYS	
Antibody Specificity	8
Antibody Specificity (Using Denatured Arrays)	11
Serum Profiling	14
RNA Binding	17
DNA Binding	19
Protein Binding	21
<hr/>	
APPENDIX	
Recipes	24
HuProt™ Literature Citations	25

Jung JG *et al.* (2014) Notch3 Interactome Analysis Identified WWP2 as a Negative Regulator of Notch3 Signaling in Ovarian Cancer. *PLoS Genet* 2014 Oct 30;**10**(10):e1004751. doi: 10.1371/journal.pgen.1004751. eCollection 2014.

Li H *et al.* (2016) Penetrance of Congenital Heart Disease in a Mouse Model of Down Syndrome Depends on a Trisomic Potentiator of a Disomic Modifier. *Genetics* 2016 Mar 30. pii: **genetics.116.188045**. [Epub ahead of print]

Xu Z *et al.* (2017) Systematic identification of the protein substrates of UDP-GalNAc:polypeptide N-acetylgalactosaminyltransferase-T1/T2/T3 using a human proteome microarray. *Proteomics* DOI: 10.1002/pmic.201600485.

Cheng X *et al.* (2017) Proteomic identification of the oncoprotein STAT3 as a target of a novel Skp1 inhibitor. *Oncotarget* **8**(2):2681-2693.



User Guide

HuProt™
Human Proteome Microarray v3.1

Page 24

TABLE OF CONTENTS

SUMMARY	1
Storage and Handling	3
Overview - Key Steps for HuProt™ Microarray Use	4
Additional Reagents and Materials	6
<hr/>	
ASSAYS	
Antibody Specificity	8
Antibody Specificity (Using Denatured Arrays)	11
Serum Profiling	14
RNA Binding	17
DNA Binding	19
Protein Binding	21
<hr/>	
APPENDIX	
Recipes	24
HuProt™ Literature Citations	25

General Appendix

Buffers:

- TBS Buffer Recipe (1X):
 - 20 mM Tris-Cl, pH 7.5
 - 150 mM NaCl
- 10X TBS:
 - For 1 liter of 10X TBS stock buffer, dissolve the following in 800 ml of distilled water:
 - 24.0 g Tris
 - 87.6 g NaCl
 - Adjust pH to 7.5 and add distilled water to 1L. Sterilize by autoclaving or by filtration, and store at room temperature.
- TBS-T (1X):
 - Add 1.0 ml Tween-20 to 100 ml 10X TBS solution. Make up to 1 liter with distilled water. Stir until homogenous (Tween-20 is very viscous and may stick to pipette tips. A 10% solution is easier to dispense than the undiluted form).
- Phosphate Buffered Saline (PBS) Recipe (1X):
 - 137 mM NaCl
 - 2.7 mM KCl
 - 10 mM Na₂HPO₄ • 2H₂O
 - 1.8 mM KH₂PO₄
- Blocking Solution Recipes:

NOTE: Blocking conditions vary depending on the protocol used. Please refer to each section for the correct recipes.



User Guide

HuProt™
Human Proteome Microarray v3.1

Page 25

TABLE OF CONTENTS

SUMMARY	1
Storage and Handling	3
Overview - Key Steps for HuProt™ Microarray Use	4
Additional Reagents and Materials	6
<hr/>	
ASSAYS	
Antibody Specificity	8
Antibody Specificity [Using Denatured Arrays]	11
Serum Profiling	14
RNA Binding	17
DNA Binding	19
Protein Binding	21
<hr/>	
APPENDIX	
Recipes	24
HuProt™ Literature Citations	25

HuProt™ Human Proteome Microarray Literature Citations

Tarrant MK *et al.* (2012). Regulation of CK2 by phosphorylation and O-GlcNAcylation revealed by semisynthesis. *Nat Chem Biol* **8(3)**:262-9.

Jeong JS *et al.* (2012). Rapid identification of monospecific monoclonal antibodies using a human proteome microarray. *Mol Cell Proteomics* **11(6)**:O111.016253.

Huang Y. *et al.* (2012). Global tumor protein p53/p63 interactome: making a case for cisplatin chemoresistance. *Cell Cycle* **11(12)**:2367-79.

Hu CJ *et al.* (2012). Identification of new autoantigens for primary biliary cirrhosis using human proteome microarrays. *Mol Cell Proteomics* **11(9)**:669-80.

Barry G *et al.* (2013). The long non-coding RNA Gomafu is acutely regulated in response to neuronal activation and involved in schizophrenia-associated alternative splicing. *Molec Psychiatry* Apr 30. doi: 10.1038/mp.2013.45. ePub ahead of print.

Lee YI *et al.* (2013). Protein microarray characterization of the S-nitrosoproteome. *Mol Cell Proteomics* **13**:63-72.

Chen Y *et al.* (2013). Bcl2-associated Athanogene 3 Interactome Analysis Reveals a New Role in Modulating Proteasome Activity. *Mol Cell Proteomics* **12(10)**:2804-19.

Donnelly CJ *et al.* (2013). RNA toxicity from the ALS/FTD C9ORF72 expansion is mitigated by antisense intervention. *Neuron* **80(2)**:415-28.

Fan B *et al.* (2013). A human proteome microarray identifies that the heterogeneous nuclear ribonucleoprotein K (hnRNP K) recognizes the 5' terminal sequence of the hepatitis C virus RNA. *Mol Cell Proteomics*. **13(1)**:84-92.

Fan, Q *et al.* (2014). Identification of proteins that interact with alpha A-crystallin using a human proteome microarray. *Mol Vis* **20**:117-124.

Deng, RP *et al.* (2014). Global Identification of O-GlcNAc Transferase (OGT) Interactors by a Human Proteome Microarray and the Construction of an OGT Interactome. *Proteomics* **14(9)**:1020-30. doi: 10.1002/pmic.201300144. ePub 2014 Mar 25.

Ma, TM *et al.* (2014). Serine Racemase Regulated by Binding to Stargazin and PSD-95: Potential NMDA-AMPA Glutamate Neurotransmission Cross-talk. *J Biol Chem* pii: jbc.M114.571604. [Epub ahead of print]; First Published on August 27, 2014.

Jung JG *et al.* (2014). Notch3 Interactome Analysis Identified WWP2 as a Negative Regulator of Notch3 Signaling in Ovarian Cancer. *PLoS Genet* 2014 Oct 30;**10(10)**:e1004751. doi: 10.1371/journal.pgen.1004751. eCollection 2014.

Liu S *et al.* (2015) Characterization of monoclonal antibody's binding kinetics using oblique-incidence reflectivity difference approach. *MAbs* **7(1)**:110-9. doi: 10.4161/19420862.2014.985919.

Cox E *et al.* (2015) Identification of SUMO E3 ligase-specific substrates using the HuProt human proteome microarray. *Methods Mol Biol* **1295**:455-63.

Hu C *et al.* (2015) Autoantibody Profiling on Human Proteome Microarray for Biomarker Discovery in Cerebrospinal Fluid and Sera of Neuropsychiatric Lupus. *PLoS One* **10(5)**: e0126643.

Hu J. *et al.* (2015) Systematic Prediction of Scaffold Proteins Reveals New Design Principles in Scaffold-Mediated Signal Transduction. *PLoS Comput Biol* **11(9)**: e1004508.

Zhang HN *et al.* (2015). Systematic identification of arsenic-binding proteins reveals that hexokinase-2 is inhibited by arsenic. *Proc Natl Acad Sci USA* **112(49)**:15084-9.

Syed P *et al.* (2015) Autoantibody Profiling of Glioma Serum Samples to Identify Biomarkers Using Human Proteome Arrays. *Sci Rep* 2015 Sep **15**:5:13895. doi: 10.1038/srep13895.

Ainscough JS *et al.* (2015) Interleukin-1β Processing Is Dependent on a Calcium-mediated Interaction with Calmodulin. *J Biol Chem* **290**:31151-31161.

Yang L *et al.* (2016) Identification of serum biomarkers for gastric cancer diagnosis using a human proteome microarray. *Mol Cell Proteomics* **15(2)**:614-23. doi: 10.1074/mcp.M115.051250. Epub 2015 Nov 23

Li H *et al.* (2016) Penetrance of Congenital Heart Disease in a Mouse Model of Down Syndrome Depends on a Trisomic Potentiator of a Disomic Modifier. *Genetics* 2016 Mar 30. pii: **genetics.116.188045**. [Epub ahead of print]

Shi L *et al.* (2016). Application of high-throughput protein array in clinical screening for tumor markers. *Int J Clin Exp Med* **9(5)**:8529-8535.

Iyama S *et al.* (2016) Drebrin: A new oncofetal biomarker associated with prognosis of lung adenocarcinoma. *Lung Cancer* **102**:74-81; doi: 10.1016/j.lungcan.2016.10.013.

Gupta S *et al.* (2017) Evaluation of autoantibody signatures in meningioma patients using human proteome arrays. *Oncotarget* 2017 Apr 10. doi: 10.18632/oncotarget.16997. [Epub ahead of print]

Chung, BK *et al.* (2017) Phenotyping and auto-antibody production by liver-infiltrating B cells in primary sclerosing cholangitis and primary biliary cholangitis. *J Autoimmunity* **77**:45-54.

Xu Z *et al.* (2017) Systematic identification of the protein substrates of UDP-GalNAc:polypeptide N-acetylgalactosaminyltransferase-T1/T2/T3 using a human proteome microarray. *Proteomics* DOI:10.1002/pmic.201600485

Cheng X *et al.* (2017) Proteomic identification of the oncoprotein STAT3 as a target of a novel Skp1 inhibitor. *Oncotarget* **8(2)**: 2681-2693.

Gupta S *et al.* (2017). Serum Profiling for Identification of Autoantibody Signatures in Diseases Using Protein Microarrays. *Methods Mol Biol* **1619**:303-315. doi:10.1007/978-1-4939-7057-5_21.

Guang, S. *et al.* (2017) Multiplexed biomarker panels discriminate Zika and Dengue virus infection in humans. *Mol Cell Proteomics* **17(2)**:349.

Guo, G *et al.* (2018) The cytomegalovirus protein US31 induces inflammation through mono-macrophages in systemic lupus erythematosus by promoting NF-κB2 activation. *Cell Death Dis* **9**:104 doi:10.1038/s41419-017-0122

Coll, JM (2018) Herpesvirus Infection Induces both Specific and Heterologous Antiviral Antibodies in Carp. *Front Immunol* **9**:39.

Bremer HD *et al.* (2018) ILF2 and ILF3 are autoantigens in canine systemic autoimmune disease. *Scientific Reports* **8**: Article number: 4852 doi:10.1038/s41598-018-23034-w





P.O. Box 569
Mayaguez, PR 00681-0569

939-280-5293
844-539-6296 (Toll Free)

www.cdi-lab.com